LARYNGOSCOPE

Equipment Details		Person Responsible for the Equipment	
Brand/Model	VZN	**Full Name**	Lecturer Merve KIRŞAN BÜYÜKTARAKÇI
Inventory Registration No		**Telephone**	0 (378) 223 53 74
Equipment Status/Quantity	Aktive/ 1 unit	**Email**	mkirsan@bartin.edu.tr
Laboratory Location	Faculty of Health Sciences / Simulation Laboratory		

Images of the Equipment



Technical Specifications

- ✓ Light is transmitted through fibre-optic bundles, providing illumination near the distal end of the laryngoscope blade.
- ✓ The blades have a one-piece construction.
- ✓ The blades attach using a hook-on fitting that locks into the designated slot on the handle.
- ✓ The light switches on automatically when the blade is attached to the handle and switches off when the blade is removed.
- ✓ The handle contains a 2.5 V lamp that provides an illumination intensity of 4,200 lux. No tools are required to replace the lamp.
- ✓ The fibre-optic light guide is securely mounted and cannot be removed without loosening the retaining screw. It can be replaced or cleaned using a screwdriver when necessary.

Adult Laryngoscope Set Supplied with the Equipment

- ✓ Three blades of different sizes
- ✓ Protective carrying case