

Information on Devices Located in the Laboratory			
Device Specifications		Device Supervisor	
Brand / Model	Laryngoscope (LOFNER)	Name Surname	Asst. Prof. Necmiye ÇÖMLEKÇİ Res. Asst.Meryem YÜCEL
Fixed Asset Registration Number	253.3.5/20/2886	Phone	0 378 501 10 00
Device Status / Quantity	1 piece	E-Mail	ncomlekci@bartin.edu.tr meryemyucel@bartin.edu.tr
Laboratory Where It Is Located	First Aid Laboratory		

Photo of the Device



Technical Specifications:

- The light of the laryngoscope is carried by fiberoptic fibers, providing illumination from the part near the tip.
- The blades are one-piece.
- The blades are attached as hook-type and there is a ball system that fits into the special slot on the handle.
- When the blade seats on the handle, the illumination bulb automatically turns on and when the blade is removed, it automatically turns off.
- There is a 4200 lux, 2.5 V bulb on the handle and no tool is needed to replace the bulb.
- The fiber optic light path is mounted in such a way that it cannot come out in any way without unscrewing the connecting screw, and when necessary, it can be replaced or cleaned with the help of a screwdriver.
- Adult Laryngoscope Set included with the device: 4 Macintosh Type Blades in different sizes, Protection and Carrying Case is included.